

Tristel Trio™

Wipes System

COMPATIBILITY

TRISTEL TRIO WIPE SYSTEM MATERIAL COMPATIBILITY WITH ULTRASOUND MEDICAL DEVICES

Manufacturer	Instrument Type	Instrument Series	Model
Advanced Tactile Imaging	Endocavity Transducer	No Series Specified	Vaginal Tactile Imager (VTI)
Bard	Ultrasound Unit	No Series Specified	Site~Rite 8
	Skin Surface Transducer	No Series Specified	32MM Linear Probe and 20MM Linear Probe
BK Medical / BK Ultrasound	Skin Surface Transducer	BK3000/BK5000	6C2, 6C2s, 9C2, 14L3, 13L4w, 10L2w, 18L5, 8L2
	Endocavity Transducer		E13C2, E14C4t, E14C4, E14 CL4b, E10C4, 20R3, X14L4
	Intraoperative Transducer		I14C5l, I14C5T, I12C5b, I12C4f
			N13C5, N11C5s, X18L5s, X12C4
	Skin Surface Transducer		5P1
	Ultrasound Transducer	Flex Focus & Pro Focus	8809, 8814, 8815, 8816, 8824, 8826, 8836, 8862, 8863, 2050, 2052, 8808, 8808e, 8818, 8838, 8848
Carestream	Skin Surface Transducer	Touch Prime	6C2, 6C2s, 9C2, 5P1, N13C5, 8L2
	Intraoperative Transducer		18L5s, 18L5, 14L3, 13L4w, 10L2w
	Endocavity Transducer		E14C4t, E14C4, E14CL4b, E10C4
Clarius Mobile Health	Handheld Ultrasound Scanners	No Series Specified	Clarius Scanner
			Arthrex Synergy MSK
Echosens	FibroScan and Ultrasound Unit	FibroScan 520 TOUCH	M6.0, D5-1268, M3.5, D7-1257, M2.5, D10-1269
Esaote	Endocavity Transducer	No Series Specified	EC123*, EC1123*
	Intraoperative Transducer	No Series Specified	IOE323*
	TOE Transducer	No Series Specified	TEE022*, TEE122*
Exact Imaging	Endocavity Transducer	ExactVu	EV9C, EV29L
Fujifilm	Endocavity Transducer	Ultrasonic Mini Probe System	PB-2020 M
GE Healthcare	Ultrasound Transducer	Logiq Systems - E9	C1-5-D, C1-6-D, C2-9-D, M6C-D, IC5-9-D, 9L-D, 11L-D, ML6-15-D, L8-18i-D, 6s-D, 6s-D, RAB2-5-D, RAB4-8-D, RIC5-9-D
		Logiq Systems - S8	C1-5-D, IC5-9-D, 9L-D, ML6-15-D
		Logiq Systems - S7	C1-5-D, 8C, IC5-9-D, 9L-D, 11L-D, ML6-15, L8-18i-D, RAB4-8-D
		Logiq P Series	4C, 5CS, 8C, BE9CS, E8C, E8CS, 8L, 9L, 11L, ML6-15, I739, T739, 3S, 5S, 6S, 7S, 4D3C-L, 4D5C-L, 4DE7C
		Logiq C Series	C2-5-RC, E8CS-RC, L6-12-RC, 3S-RC
		Logiq A Series	3.5CS, 5CS, 8C, BE9CS, E8CS, 8L, 9L, 10L, 11L, 12L, I739, T739, 3S, 5S, 7S
		Logiq E Series	4C-RS, 8C-RS, E8C-RS, 8L-RS, 9L-RS, 12L-RS, i12L-RS, L8-18i-RS, I739-RS, T739-RS, 3S-RS, 6S-RS
		Vivid E9	4C-D, 9L-D, 11L-D, ML6-15-D, 6S-D, 12S-D, 4V-D
		Vivid P Series	4C, 5CS, 8C, E8CS, 8L, 11L, T739, 3S, 6S
		Vivid S Series	4C-RS, 8C-RS, E8C-RS, 8L-RS, 9L-RS, 12L-RS, 3S-RS, 3Sc-RS, 5S-RS, 6S-RS, 7S-RS, 12S-RS, M4S-RS

TRISTEL TRIO WIPES SYSTEM MATERIAL COMPATIBILITY WITH ULTRASOUND MEDICAL DEVICES

Manufacturer	Instrument Type	Instrument Series	Model
GE Healthcare	Ultrasound Transducer	Vivid e/i/q Series	4C-RS, 8C-RS, E8C-RS, 8L-RS, 9L-RS, 12L-RS, 3S-RS, 3Sc-RS, 5S-RS, 6S-RS, 7S-RS, 12S-RS, M4S-RS
		Voluson Expert Series	AB2-7-D, 4C-D, C1-5-D, C4-8-D, M6C-D, IC5-9-D, 9L-D, 11L-D, ML6-15-D, 3S-D, RAB2-5-D, RAB4-8-D, RIC5-9-D, RIC6-12-D, RRE5-10-D
		Voluson Performance Series	4C-RS, E8C-RS, 12L-RS, 3Sc-RS, RAB2-5-RS, RIC5-9W-RS
		Voluson Signature Series	4C-RS, AB2-7-RS, C1-5-RS, 8C-RS, E8C-RS, 9L-RS, 12L-RS, ML6-15-RS, 3Sc-RS, RAB2-5-RS, RAB4-8-RS, RIC5-9W-RS
		Venue 40 Series	4C-SC, E8CS-SC, 12L-SC, L8-18i-SC, 3S-SC
	TOE Transducer	No Series Specified	6T, 6T-RS, 6Tc, 6Tc-RS, 6Tv, 9T, 9T-RS, 6VT-D
Hitachi	Ultrasound Transducer	No Series Specified	C22I, C22K, C22P, C22T, C25, C25I, C252, C25P, C35, C41, C41B, C41L47RP, C41RP, C41V, C41V1, C42, C42K, C42T, CC41R, CC41R1, CL4416R, L34, L43K, L44, L44I, L44K, L44LA1, L46K, L46K1, L51K, L53K, L53K1, L55, L64, S12, S121, S21, S211, S22, S31, S31KP, S42, SML44, VC41V, ASU-1000B-3.5, ASU-1000C-3.5, ASU-1001, ASU-1002, ASU-1003, ASU-1005, ASU-1009, ASU-1010, ASU-1011, ASU-1012, ASU-1013, ASU-35-3, ASU-35CWD-2, ASU-35WL-10, ASU-35WSJ-7.5, ASU-36WL-10, ASU-64B, ASU-65B, ASU-67, UST-2265-2, UST-2266-5, UST-5024N-3.5, UST-5037P-3.5, UST-5039P-3.5, UST-5043P-3.5, UST-5044-3.5, UST-5045P-3.5, UST-5046-3.5, UST-5047-5, UST-52103, UST-52105, UST-52108, UST-52109, UST-52114P, UST-52124, UST-52127, UST-52129, UST-5226D-3.5/2, UST-5236D-2.5/2, UST-5242D-2.5/2, UST-5244-3.5, UST-5255-2.5, UST-5256-3.5, UST-5257-5, UST-5263D-2.5/2, UST-5265-2.5, UST-5266-3.5, UST-5267-5, UST-5268P-5, UST-5274-2.5, UST-5275-3.5, UST-5276-5, UST-5279-3.5, UST-5281-5, UST-5283-2.5, UST-5284-2.5, UST-5285-3.5, UST-5286-2.5, UST-5287-3.5, UST-5292-5, UST-5294-5, UST-5295, UST-5296, UST-5297, UST-5298, UST-5299, UST-533, UST-534, UST-535, UST-536, UST-5410, UST-5410H, UST-5411, UST-5412, UST-5413, UST-5415, UST-5417, UST-5419, UST-547, UST-5512U-7.5, UST-5514DTU-7.5, UST-5515-7.5, UST-5521L-7.5, UST-5523L-7.5, UST-5524-5, UST-5524-7.5, UST-5526L-7.5, UST-5531, UST-5534T-7.5, UST-5539-7.5, UST-5540P-7.5, UST-5542, UST-5543, UST-5545, UST-5546, UST-5548, UST-5551, UST-5561-7.5, UST-556T-7.5, UST-556TU-7.5, UST-567, UST-568, UST-5710-7.5, UST-5711-7.5, UST-5712, UST-5713T, UST-574T-7.5, UST-575-7.5, UST-576T-7.5, UST-577-7.5, UST-579T-7.5, UST-5818-5, UST-5819T-5, UST-657-5, UST-660-7.5, UST-665P-5, UST-667P-5, UST-670P-5, UST-675P, UST-676P, UST-677P, UST-9101-7.5, UST-9102-3.5, UST-9102U-3.5, UST-9103-5, UST-9104-5, UST-9111-5, UST-9112-5, UST-9113P-3.5, UST-9114-3.5, UST-9115-5, UST-9116P-5, UST-9118, UST-9119, UST-9120, UST-9121, UST-9123, UST-9124, UST-9125-7.5, UST-9126, UST-9127, UST-9128, UST-9130, UST-9132I, UST-9132T, UST-9133, UST-9135P, UST-9136, UST-9136U, UST-9137, UST-9145, UST-9146I, UST-9146T, UST-9147, UST-9148, UST-9149, UST-9151, UST-932-5, UST-932-7.5, UST-935N-5, UST-939-3.5, UST-939D-3.5, UST-943-3.5, UST-944B-3.5, UST-945B-5, UST-945BP-5, UST-952-5, UST-953P-5, UST-954-5, UST-957-7.5, UST-961-5, UST-961-7.5, UST-964P-5, UST-968P-5, UST-969-5, UST-970-7.5, UST-972-3.5, UST-974-5, UST-975-7.5, UST-977-5, UST-978-3.5, UST-979-3.5, UST-980-5, UST-980P-5, UST-981-5, UST-981P-5, UST-984-5, UST-984P-5, UST-985-7.5, UST-987-7.5,

TRISTEL TRIO WIPES SYSTEM MATERIAL COMPATIBILITY WITH ULTRASOUND MEDICAL DEVICES

Manufacturer	Instrument Type	Instrument Series	Model
Hitachi	Ultrasound Transducer	No Series Specified	UST-988-2.5, UST-989-3.5, UST-989D-3.5, UST-990-5, UST-992-5, UST-994P-5, UST-995-7.5, UST-996P-5, UST-MC11-8731, EUP-B314, EUP-B512, EUP-B514, EUP-B712, EUP-B715, EUP-C321T, EUP-C511, EUP-C514, EUP-C516, EUP-C524, EUP-C532, EUP-C715, EUP-CV714, EUP-F334, EUP-F531, EUP-L34T, EUP-L52, EUP-L53, EUP-L53L, EUP-L53S, EUP-L54MA, EUP-L73S, EUP-L74M, EUP-L75, EUP-LS85, EUP-O33JA, EUP-O53T, EUP-O54J, EUP-O732T, EUP-S50A, EUP-S52, EUP-S63, EUP-S70, EUP-S72, EUP-S80, EUP-V53W, EUP-V73W, EUP-VV731
Mcube	Bladder Scanner	Cubescan	BioCon-700 and BioCon-900
Medcom	Endocavity Transducer	BiopSee Bi Plane Transrectal Transducer	BIPL7.5/70/128, BIC6.5/10/128
Mindray	Ultrasound Transducer	No Series Specified	C5-1S, C5-1E, C5-1U, C5-2, C5-2S, C5-2E, C6-2E, C6-2GS, C6-2GE, C7-3E, C11-3E, C11-3S, D7-2, D7-2E, D7-2S, D6-2, D6-2P, L7-3, L7-3S, L7-3E, L9-3U, L11-3U, L11-4, L11-4S, L12-3, L12-3E, L12-3S, L12-4, L12-4S, L14-6, L14-6S, L14-6N, L14-6NS, L14-6NP, L14-6NE, L14-6WE, L14-6WS, L14-6WU, LM14-6E, LM14-6S, LM16-4U, P4-2, P4-2S, P4-2E, P4-2NE, P10-4E, P10-4S, SC6-1U, SP5-1U, SP5-1S, SP5-1E, V11-3HU, V11-3U, 3C1, 3C1S, 3C1P, 10L24EA, 35C50EA, 35C50EB, 35C50EA, 35C50P, 65C15EAV, 65C15EA, 75L38EA, 75L38EB, 75L53EA, 75L60EA, 75L38P
	TOE Transducer	No Series Specified	P7-3T, P7-3TE
Philips*	Ultrasound Transducer	iU22	C8-4v (p/n 4535611 7292x where x is 3 or greater and p/n 4535612 8750x where x is 2 or greater)
		HD15	C8-4v (p/n 4535611 7292x where x is 3 or greater and p/n 4535612 8750x where x is 2 or greater)
		HD11 XE	C8-4v (p/n 4535611 7292x where x is 3 or greater and p/n 4535612 8750x where x is 2 or greater)
		HD9	C8-4v (p/n 4535611 7292x where x is 3 or greater and p/n 4535612 8750x where x is 2 or greater)
		HD7 XE	C8-4v (p/n 4535611 7292x where x is 3 or greater and p/n 4535612 8750x where x is 2 or greater)
		HD7	C8-4v (p/n 4535611 7292x where x is 3 or greater and p/n 4535612 8750x where x is 2 or greater)
		HD6	C8-4v (p/n 4535611 7292x where x is 3 or greater and p/n 4535612 8750x where x is 2 or greater)
		HD5	C8-4v (p/n 4535611 7292x where x is 3 or greater and p/n 4535612 8750x where x is 2 or greater)
	Endocavity Transducer	CX50	C9-3v
		Sparq	
		CX30	
		EPIQ	
		HD15	BP10-5ec
		HD11 XE	
		EPIQ	
		Affinity 50	
		Affinity 70	C9-4v
		EPIQ7	
		EPIQ5	
		Affinity 50	
		Affinity 70	
		CX50	

TRISTEL TRIO WIPES SYSTEM MATERIAL COMPATIBILITY WITH ULTRASOUND MEDICAL DEVICES

Manufacturer	Instrument Type	Instrument Series	Model
Philips*	Endocavity Transducer	EPIQ7	C10-3v
		EPIQ5	
		iU22	
		Affinity 70	
		CX50	
	Skin Surface Transducer	EPIQ7	S5-1
		EPIQ5	
		Affinity70	
		CX50	
		iU22	
		iE33	S12-4
		Sparq	
		EPIQ7	
		EPIQ5	
		Affinity70	
		Affinity50	
		iE33	
		HD5	
		HD15	
		HD11	
		CX50	
		HD11 XE	
		No Series Specified	3D6-2
			3D8-4
	Endocavity Transducer	HD11	3D9-3v
		HD15	
		iU22	
		ClearVue	
		EPIQ	
		Affinity	7.5 MHz
		iU22	
		Sparq	C8-4v
		Affinity	
	Skin Surface Transducer	CX30	C6-2 Compact
		Sparq	
		Affinity	
		Sparq	C6-3

TRISTEL TRIO WIPES SYSTEM MATERIAL COMPATIBILITY WITH ULTRASOUND MEDICAL DEVICES

Manufacturer	Instrument Type	Instrument Series	Model
Philips*	Skin Surface Transducer	CX30	C8-5 Compact
		CX50	
		Affinity	
		EPIQ	
		iE33/iU22	X3-1
		EPIQ7/iE33	X5-1
		EPIQ7/iE33/iU22	X7-2
		CX50	C5-1
		EPIQ7	
		QPIQ5	
		Affinity 70	
		iU22	
		iE33	
		HD15	
		ClearVue	C5-2
		Lumify	
		Xperius	
		iE33	
		iU22	
		HD15	
		HD11 XE	
		HD11	
		HD9	
		HD7 XE	
		HD7	
		HD6	
		HD5	
		EPIQ	C9-2
		CX50	C9-4
		No Series Specified	C9-3io
	Endocavity Transducer	iE33	C9-4v
		EPIQ7	
		EPIQ5	
		Affinity50	
		Affinity70	
		CX50	
		CX50	C9-4ec
		EPIQ	C9-4v Compact

TRISTEL TRIO WIPES SYSTEM MATERIAL COMPATIBILITY WITH ULTRASOUND MEDICAL DEVICES

Manufacturer	Instrument Type	Instrument Series	Model
Philips*	Endocavity Transducer	CX50	C9-5ec
	Skin Surface Transducer	EPIQ7	C8-5
		EPIQ5	
		Affinity 70	
		Affinity 50	
		iE33	
		iU22	
		HD7 XE	
		HD11 XE	
		HD9	
		HD11	
		HD15	
		CX30	
		CX50	
		Sparq	
	Endocavity Transducer	EPIQ	C10-4ec
	Skin Surface Transducer	iE33	L8-4
		iU22	
		HD11 XE	
		CX50	
		iE33	L9-3
		iU22	
		HD15	
		HD9	
		HD11 XE	
		CX50	
		No Series Specified	L10-4lap
		CX50	L11-3
		iE33	

TRISTEL TRIO WIPES SYSTEM MATERIAL COMPATIBILITY WITH ULTRASOUND MEDICAL DEVICES

Manufacturer	Instrument Type	Instrument Series	Model
Philips*	Skin Surface Transducer	HD5	L12-3
		HD6	
		HD7	
		HD11	
		HD15	
		Affinity	
		CX50	
		EPIQ	
		Lumify	L12-4
		Xperius	
		ClearVue	
		CX30	L12-4 Compact
		Sparq	
		CX50	
		Affinity	
		Sparq	L12-5
		CX50	L12-5 Compact
		Affinity	
		EPIQ	
		iU22	L17-5
		iU22	L18-5
		Sparq	L15-7IO
		EPIQ7	
		EPIQ5	
		Affinity 70	
		Affinity 50	
		iE33	
		iU22	
		HD11	
		HD15	
		CX30	
		CX50	
	TOE Transducer	iE33	T6207

TRISTEL TRIO WIPES SYSTEM MATERIAL COMPATIBILITY WITH ULTRASOUND MEDICAL DEVICES

Manufacturer	Instrument Type	Instrument Series	Model
Philips*	TOE Transducer	iE33	D2cwc
		EPIQ7	
		EPIQ5	
		iU22	
		CX50	
		HD11	
		HD9	
		HD15	
		HD11 XE	
	Ultrasound Transducer	No Series Specified	D5cwc
	TOE Transducer	No Series Specified	X8-2t, X8-2ti
		EPIQ7	D2tcd
		EPIQ5	
		iE33	
		iU22	
		HD11 XE	
		EPIQ7	S8-3
		EPIQ5	
		Affinity50	
		Affinity70	
		CX30	
		CX50	
		iE33	
		iU22	
		HD11	
		HD11 XE	
		HD15	
	Skin Surface Transducer	iU22	S3-1
		iU22	S4-1
		ClearVue	
		Lumify	
		iU22	S4-2
		CX30	S4-2 Compact
		Sparq	
		Affinity	
		Affinity	S5-1
		EPIQ	

TRISTEL TRIO WIPES SYSTEM MATERIAL COMPATIBILITY WITH ULTRASOUND MEDICAL DEVICES

Manufacturer	Instrument Type	Instrument Series	Model
Philips*	Skin Surface Transducer	iU22	S5-2
		iU22	S8
		CX30	S8-3 Compact
		CX50	
		HD5	
		Affinity	
		EPIQ	
	TOE Transducer	No Series Specified	S8-3t
			S6-2mpt
		iE33	S7-2t
		iU22	
		HD15	
		HD11 XE	
		HD11	
		HD9	
		HD7 XE	
		HD7	
		HD6	
		HD5	
		iE33	S7-2 Omni (21778A)
		iU22	
		HD15	
		HD11 XE	
		HD11	
		HD9	
		HD7 XE	
		HD7	
		HD6	
		HD5	
		No Series Specified	S7-3t
		HD5	S12
	Skin Surface Transducer	Affinity	S12-4 Compact
		EPIQ	
	TOE Transducer	iE33	T6H/21378A OMNI III
		HD11 XE	
		HD7 XE	

TRISTEL TRIO WIPES SYSTEM MATERIAL COMPATIBILITY WITH ULTRASOUND MEDICAL DEVICES

Manufacturer	Instrument Type	Instrument Series	Model
Philips*	Skin Surface Transducer	HD11 XE	V6-2
		HD15	
		HD5	
		iU22	
		Affinity	
		ClearVue	
		EPIQ	
		iU22	V8-4
		iU22	VL13-5
		EPIQ	X6-1
	Pencil Transducer	No Series Specified	1.9MHz CW
			5.0MHz CW
			5.0MHz CW Pencil
	TOE Transducer	EPIQ7	X7-2t
		EPIQ5	
		Affinity70	
		iE33	
		CX50	
		Sparq	
Samsung*	Ultrasound Transducer	No Series Specified	C2-8ED
	Skin Surface Transducer		CA1-7A, C2-6IC, CA2-8A, CF2-8, CA2-8AD, L5-13IS, L7-16IS, LA3-16AI, L3-12A, LA3-14AD, L5-13/50, CV1-8A, CV1-8AD, 3DC-6EN, CA2-9A, CA2-9AD, CA1-7AD, CA2-6BM, L3-8ED, LS6-15, LA3-14AD, LA3-14A, PA1-5A, PA1-5AD, SP3-8, CA3-10A
	Endocavity Transducer		ER4-9/10ED, EV4-9/10ED, V5-9, 3D4-9ES, EVN4-9/10ED, 3D4-9ES, EV3-10B, EVN4-9, EC4-9CD, VR5-9, E3-12A, EA2-11B
Siemens*	Skin Surface Transducer	Acuson Antares	CH4-1, CH6-2, CW2, PH4-1
	Endocavity Transducer		EC9-4
	Skin Surface Transducer	Acuson CV70	3V2c, 7V3c, CW2, CW5
		Acuson Cypress	3V2c, 4C1, 7V3c, Aux CW
		Acuson S1000	4V1c, 4C1, 4V1, 6C2, 8V3, CW2, CW5
	Endocavity Transducer		EC9-4, EV-8C4
	Skin Surface Transducer	Acuson S2000	4V1c, 4C1, 4V1, 4V1c, 6C2, 8V3, CW2, CW5
	Endocavity Transducer		EC9-4, EV-8C4
	Ultrasound Transducer	Acuson SC2000 Prime	4V1c, 8V3, AUX CW, CW2
		Acuson S3000	4C1, 4V1, 4V1c, 6C2, 8V3, CW2, CW5, EC9-4, EV-8C4
		Acuson SC2000	4V1c, 8V3, CW2

TRISTEL TRIO WIPES SYSTEM MATERIAL COMPATIBILITY WITH ULTRASOUND MEDICAL DEVICES

Manufacturer	Instrument Type	Instrument Series	Model
Siemens*	Ultrasound Transducer	Acuson Sequoia	EV-8C4
	Skin Surface Transducer		15L8, 3V2C, 4C1, 4V1/4V1c, 5V2c, 6C2, 6L3, 7V3c, 8C4, 8L5, 8L5T, 8V3c, 8V5, Aux CW
	Endocavity Transducer		EC-10C5
	Ultrasound Transducer	Acuson X700	4C1, 4V1c, 6C2, CW2, CW5, P8-4, VF16-5
	Ultrasound Transducer	Acuson X150	EV9-4
	Skin Surface Transducer		CH5-2, P4-2, P8-4
	Endocavity Transducer		EC9-4
	Ultrasound Transducer	Acuson X300	CH5-2, CW2, CW5, EC9-4, EV9-4, P4-2, P8-4
		Acuson X300PE	CH5-2, CW2, CW5, EC9-4, EV9-4, P4-2, P8-4
		Acuson X500	CW2, CW5, EC9-4
		Sonoline G20	EC9-4
		Sonoline G40	CH5-2, EC9-4, EV9-4, P4-2
		Sonoline G50/G60	CW2, CW5, EC9-4
	TOE Transducer	Acuson Cypress	V5Ms (SN 327xxxxx & 328xxxxx)
		Acuson S1000	V5Ms (SN 327xxxxx & 328xxxxx)
		Acuson S2000	V5Ms (SN 327xxxxx & 328xxxxx)
		Acuson S3000	V5Ms (SN 327xxxxx & 328xxxxx)
		Acuson SC2000	V5Ms (SN 327xxxxx & 328xxxxx)
		Acuson SC2000 Prime	V5Ms (-SN 327XXXXX)
		Acuson SC2000 Prime	V5Ms (SN 328XXXXX -)
		Acuson Sequoia	V5Ms (SN 327xxxxx & 328xxxxx)
		Acuson X300PE	V5Ms (SN 327xxxxx & 328xxxxx)
		Acuson X700	V5Ms (SN 327xxxxx & 328xxxxx)
		Acuson SC2000	Z6Ms
SonoScape*	Ultrasound Transducer	E30 Series (E30/ E30 Pro/ E30 Exp/ E30S/ E30 Plus/ E30VO/ E29)	VC2-9A, VC2-9, VE3-10, VC6-2, VC8-3, VC311, VE9-5, VL12-5, VC6-2, EC2, EC9-5, 5V1, 6V1, 6V3, 6V4, 6V5, BCC9-4, BCC9-5, BCL10-5, 6V-A, 6V1A, 6V2A, 6V3A, 6V7, 8V-A, 6CT4, C3-10V, 2P1, 5P1, 10L1, C311, C321, C343, C344, C351, C352, C361, C362, C511, C521, C541, C542, C543, C611, C612, C721, L541, L561V, L662V, L741, L741V, L742, L743, L745, L746, L751, L752, L761V, L744, C312, C322, C345, C353, C354, C544, C613, C5-1, 3C-A, C1-6, 2P2, 3P1, 3P-A, 4P-A, 4P-B, 5P2, 7P-A, 7P-B, 8P1, 12P-A, 12P-B, P5-1, 3C-B, 9L-A, 13L-A, 18L-A, S1-5, C1-6A, C2-9, 12L-A, 12L-B, ML3-18, C9-5

TRISTEL TRIO WIPES SYSTEM MATERIAL COMPATIBILITY WITH ULTRASOUND MEDICAL DEVICES

Manufacturer	Instrument Type	Instrument Series	Model
Sonosite Fujifilm	Portable Ultrasound System	FC1-System	Powerplug
		M-Turbo-System	M-Turbo-System
			M-Turbo-TTC
			Foot Paddle
			Powerplug
			Mini-Dock
		NanoMaxx-System	NanoMaxx-System
			Dock
			Powerplug
		M Turbo M-OB/GYN Office	(System & Dock)
Sonosite Fujifilm	Ultrasound Transducer	No Series Specified	SLT
	Skin Surface Transducer	No Series Specified	P11x
			L52/L52e - Vet Only
	Ultrasound Transducer	Edge-System	P11x2
		Edge II-System	P11x2
		S-System	P11x2
		SII-System	P11x2
		M-Turbo-System	P11x2
		No Series Specified	D2/D2x
	Ultrasound System	iViz System	iViz System
			Protective Cover
	Ultrasound Unit	Edge-System	Edge-TTC
			Powerplug
			Mini-Dock
		Edge II-System	Edge II-System
			Edge-TTC
			Powerplug
			Mini-Dock
		S-System	S-System
			S-Series-TTC
			Powerplug
		SII-System	SII-System
			Powerplug
	Ultrasound Unit System	X-Porte - System	Powerplug
	TOE Transducer	No Series Specified	TEExi Transducer

TRISTEL TRIO WIPES SYSTEM MATERIAL COMPATIBILITY WITH ULTRASOUND MEDICAL DEVICES

Manufacturer	Instrument Type	Instrument Series	Model
Canon (Toshiba)	Ultrasound Transducer	Aplio & Viamo Series	PLT-308P, PVT-770RT, PLT-705BTF/705BTH, PST25ST/50BT, PVT712BT, PST-25BT, PST-30BT, PLT-704SBT, , PLT704ST, PLT-1204ST, PVT-375ST,
	Pencil Transducer		PC-20M, PC-50M,
	TOE Transducer		PET-508MA, PET-512MA/512MC, PET511BTM
	Endocavity Transducer		PVT-661VT, PVT-781VT, PVT-781VTE, PVL-715RST, PVT-770RT, PVT-781VT, PVT-681MV,
	Skin Surface Transducer		PVT-674BT, PLT-705BTH, PVT375ST *, PVT-7350BTP, PLT-1204BX, PLT-704SBT/705BT/1005BT, PLT704ST, PLT1204ST, PVT375ST, PST-25SX, PST-65AT, PLT1204ST, PLT-1204MV, PVT-375MV/382MV/675MV/675MVL, PST25ST, PST50BT, PST-30BT, PVT-375BT, PVT-375SC, PVT-382BT, PVT712BT, PLT-604AT, PLT-704AT, PLT-805AT, PLT-1204BT, PLT-1204BX, PLT-705BT, PLT-1005BT, PVT-647BT, PST-25ST/50BT, PVT-712BT, PST-25BT/50AT/65AT
	Intraoperative Transducer		PVT-745BTF, PVT-745BTH, PLT-705BTF, PVT-745BTV, PVT-745BTF/745BTH, PVT-745BTV, PLT-1202S
	Linear Transducer		PLT-308P
	Intraoperative Transducer	Nemio & Famio Series	PVF-738F/738H/745V
			PLG-506M/805S
	Skin Surface Transducer		PVG-366M/381M/681S/720S, PLQ-703A/805A/1203A, PVQ-375A/381A, PVQ-662A/674A
	Endocavity Transducer		PVG-600S/601V/630V, PVQ-641V
	Skin Surface Transducer		PLM-503AT/703AT/805AT
			PLM-1204AT
	Intraoperative Transducer		PLM-1202S
	Skin Surface Transducer		PVF-381MT, PVF-575AT
	Endocavity Transducer		PVF-620ST
	Ultrasound Transducer		PVM-375AR
	Skin Surface Transducer		PVM-381AT
	Endocavity Transducer		PVM-620ST, PVM-621VT
			PVM-651VT
	Skin Surface Transducer		PVM-656AT, PVM-662AT, PVM-672AT
			PSM-20CT/25AT/30BT
	Ultrasound Transducer		PVM-656A
	Intraoperative Transducer		PVF-738F*, PVF-738H*, PVF-745V*, PLG-308P, PLM-1202S,
	Endocavity Transducer		PVG-600S*, PVG-601V*, PVG-630V*, PVM-651VT, PVM-740RT, PVL-715R, PVL-715RT, PVQ-641V, PVF-3620ST,
	Laparoscopic Transducer		PVM-787LA
	Skin Surface Transducer		SMA-736SA, PC-50V, PVG-366M*, PVG-381M*, PVG-681S*, PLF-308P, PLM-1204AT, PVF-381MT, PSM-37AT, PSM-70AT, PVF-575AT, PVQ-375A, PVQ-674A, PVM-375AT, PSM-37CT, PSM-50AT, PLG-506M*, PLG-805S*, PLQ-703A, PLQ-805A, PLQ-1203A, PLM-503AT, PLM-703AT, PLM-805AT, PSM-20CT, PSM-30BT
	Pencil Transducer		PC-19M

TRISTEL TRIO WIPES SYSTEM MATERIAL COMPATIBILITY WITH ULTRASOUND MEDICAL DEVICES

Manufacturer	Instrument Type	Instrument Series	Model
Canon (Toshiba)	Skin Surface Transducer	Xario Series	PSU-25BT/30BT & PVU-375BT/382BT
			PSU-50BT & PVU712BT
			PLU-704BT
			PVU-745BTV
			PVU-745BTF/745BTH/770ST
	Endocavity Transducer		PVL-715RS
	Skin Surface Transducer		PVU-674BT & PLU-1005BT, PVU-781VT/ 781VTE
	Endocavity Transducer		PVU-781VT
	Skin Surface Transducer		PLU-1204BT
	Skin Surface Transducer	Xario Series	PVU-745BTV, PVU-745BTF, PVU-745BTH, PVU-681MVL, PVU-375BT, PVU-382BT, PSU-50BT, PSU-25BT, PSU-30BT, PSU-70BT, PVU712BT, PLU-704BT
	Endocavity Transducer		PVL-715RS, PVU-770ST,
	TOE Probe	Aplio & Viamo Series	PET-508MA, PET-512MA, PET-512MC, PET511BTM

* Compatible with the Tristel Sporidical Wipe, part of the Tristel Trio Wipe System

Disclaimer: Material compatibility testing performed via different methods including wiping and soaking. Tested materials subjected to varying number of cycles.

For more information contact Tristel team on cooperation@tristel.com

Tristel

Created by: Tristel Solutions Limited, Lynx Business Park, Cambs, UK, CB8 7NY
T +44 (0) 1638 721500 - E mail@tristel.com - W www.tristel.com

Ireland: Brennan & Company, 61 Birch Avenue, Stillorgan Industrial Park, Stillorgan, Co. Dublin, A94 XW68 - T 01 2952501 - E enquiries@brennanco.ie

Australian Sponsor: Tristel Pty Ltd, 40/328 Reserve Road, Cheltenham, VIC 3192
T 1300 680 898 - F +61 (0)3 9533 6193 - E mail-au@tristel.com

For Tristel patent information please visit: <http://www.our-patents.info/tristel>

New Zealand: Tristel New Zealand Limited, 23 Birch Avenue, Judea, Tauranga P.O. Box 3110
T +64 (0)7 5771560 - F +64 (0)7 5771567 - E info@tristel.co.nz

Hong Kong & Taiwan: 21st Floor, 168 Electric Road, Fortress Hill, Hong Kong
T +852 2895 6968 - F +852 2869 4388 - E customerservice@tristel.com.hk

United Arab Emirates (UAE), Saudi Arabia, Qatar, Bahrain, Oman, Kuwait, Lebanon, Egypt, Iran, Iraq, Syria, Sudan, Yemen: Fal Care FZE, Saif desk Z2-04, PO Box 514484, Sharja, United Arab Emirates - T 00971 54 707 049 - E info@falcare.com

Israel: Sachar Medical Technologies, 20 Zvi Berkman St., Petach Tiquwa, 4927976, Israel
T 00972 544513925 - E sachar.medtech@gmail.com

Copyright © Tristel Solutions
Mkt-Val-449-4
December 2018